



MISSOURI DEPARTMENT OF HEALTH
STATE PUBLIC HEALTH LABORATORY
DATAMASTER MAINTENANCE REPORT

BREATH ALCOHOL PROGRAM

JUN - 8 2009

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send copy to Department of Health; retain original in department file.

| | |
|---|---------------------------------------|
| DATAMASTER SN BAC Datamaster 950089 | DATE OF INSPECTION 06/03/09 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 104 W. 3rd Street Lawton Mo. 64062 | TIME OF INSPECTION 1422 |

CHECKLIST: Place a check (✓) to the left of each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unchecked items must be corrected before using instrument.

DIAGNOSTIC CHECK (PRINTOUT ATTACHED)

| | |
|---|---|
| <input checked="" type="checkbox"/> COMPUTER | <input checked="" type="checkbox"/> DETECTOR |
| <input checked="" type="checkbox"/> PROGRAM | <input checked="" type="checkbox"/> FILTERS |
| <input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C | <input checked="" type="checkbox"/> QUARTZ STANDARD |
| <input checked="" type="checkbox"/> FLOW DETECTOR | <input checked="" type="checkbox"/> CALIBRATION |
| <input checked="" type="checkbox"/> PUMP HIGH SPEED | <input checked="" type="checkbox"/> PRINTER |

INDICATOR LIGHTS

TIME AND DATE

SIMULATOR TEMPERATURE (34 °C ± 0.2°C) **(34.1 °C)**

CALIBRATION CHECK -
Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) (USE RECIRCULATION PUMP)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE
(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

| | | |
|--|--|--|
| TEST 1 <input checked="" type="checkbox"/> 0.102% | TEST 2 <input checked="" type="checkbox"/> 0.102% | TEST 3 <input checked="" type="checkbox"/> 0.103% |
|--|--|--|

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS: (DO NOT INCLUDE SIMULATOR TESTS)

| | | | | | | | | | | | |
|----------|---|---------|-------------------------------------|-----------|-------------------------------------|-----------|-------------------------------------|-----------|-------------------------------------|------------|-------------------------------------|
| REFUSALS | / | (0-.04) | <input checked="" type="checkbox"/> | (.05-.09) | <input checked="" type="checkbox"/> | (.10-.14) | <input checked="" type="checkbox"/> | (.15-.19) | <input checked="" type="checkbox"/> | (Over .19) | <input checked="" type="checkbox"/> |
|----------|---|---------|-------------------------------------|-----------|-------------------------------------|-----------|-------------------------------------|-----------|-------------------------------------|------------|-------------------------------------|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary)

INSTRUMENT CONFORMS TO DOH-SS STANDARDS

| | | | |
|--|---|--------------------------|--------------------|
| MANUFACTURER: REPCO | CONCENTRATION: 0.100% | EXPIRES: 06/11/09 | LOT#: 09001 |
| INSPECTING OFFICER | | | |
| SIGNATURE P. Brian LaFaver | PRINT NAME P. Brian LaFaver | | |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 820 221 07/23/10 | TELEPHONE NUMBER Lawton P.D. (816) 580-7210 | | |

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 08001

EXPIRATION DATE: August 11, 2009 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 08001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed utilizing a gas chromatograph and found to contain 0.1216 gms/dl wt/vol ethyl alcohol in aqueous solution (i.e. ethanol).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of 0.100% +/-3% when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator.

The date of manufacture for this lot number is February 12, 2008.
The expiration date for this lot number is August 11, 2009 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Gartner, President
RepCo Marketing, Inc.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
LAWSON POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 950089
06/03/09
14:22

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
HEATERS: 50c
SAMPLE CHAMBER: OKAY
FLOW DETECTOR: OKAY
PUMP: OKAY
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefg hijklmno
pqrstuvwxyz{|}~*

OPERATOR SIGNATURE

CARD STK # 60036
REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
LAWSON POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 950089
06/03/09

TESTING OFFICER:

LAFAYOR
OFFICER I.D.: 200
PERMIT NUMBER: 820221
EXPIRATION DATE: 07/23/10
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST 14:25
INTERNAL STANDARD 14:25
EXTERNAL STANDARD 14:25
BLANK TEST 14:26
EXTERNAL STANDARD 14:26
BLANK TEST 14:27
EXTERNAL STANDARD 14:27
BLANK TEST 14:28

N = 3
SIM. = .1
AVG. = .1023

OPERATOR SIGNATURE

CARD STK # 60036
REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
LAWSON POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 950089
06/03/09

ARREST TIME: 12:00

SUBJECT NAME: LAFAYOR
DOB: 03/23/70 SEX: M
STATE/D.L.: MO/
ARRESTING OFFICER:
LAFAYOR

OFFICER I.D.: 200
TESTING OFFICER:
LAFAYOR

OFFICER I.D.: 200
PERMIT NUMBER: 820221
EXPIRATION DATE: 07/23/10
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST 14:31
INTERNAL STANDARD 14:31
RADIO INTERFERENCE .000
VERIFIED

OPERATOR SIGNATURE

CARD STK # 60036
REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II




P. BRIAN LAFAVOR

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air, tested under the provisions of sections 577.020 through 577.041, RSMo 1900.

Date 07/23/08
Number B20221
Expires 07/23/2010


Director of State Public Health Laboratory

Director, Department of Health